



[See Rule 15(3)]
 Government of West Bengal
OFFICE OF THE CONTROLLER OF LEGAL METROLOGY
Certificate of verification

SBVCA **1922932** ORIG

Name of Legal Metrology Officer: **R. PAUL, WB-262**

Unit No.: 67, Kolkata (South)

No.

I hereby certify that I have this day ^{re} verified and stamped / rejected the under mentioned weights, measures, etc.

Belonging to

Deys Medical Stores (MFG.) Ltd.

Locality *62, Bunder Road, PO-13, P.S.-Gariah*

Quantity	Denomination		Weighing instrument				Measuring instrument	Verification Fee		Carriage Conveyance adjusting charges etc.
	Weights	Measures	Capacity	Class	Manufacturer	Type		Rs.	P.	
1	2 Max	3 Min	4 L	5	6	Model	7/8	9	10	
<i>NT/12</i>	<i>1 kg</i>	<i>250 mg</i>	<i>10 mg</i>	<i>II</i>	<i>Gold Tech.</i>	<i>GTA</i>	<i>20103292</i>	<i>4890.00/-</i>	<i>upn 76</i>	
	<i>200 kg</i>	<i>400 g</i>	<i>20 g</i>	<i>III</i>	<i>GTEP Goldtech</i>	<i>GTEP</i>	<i>19081877</i>			
	<i>50/100 kg</i>	<i>100 g</i>	<i>5/10 g</i>	<i>III</i>	<i>Goldtech</i>	<i>GTPD</i>	<i>19A066459</i>			
	<i>5 kg</i>	<i>10 g</i>	<i>0.5 g</i>	<i>III</i>	<i>Goldtech</i>	<i>GTET</i>	<i>40374499</i>			
	<i>420 g</i>	<i>200 mg</i>	<i>10 mg</i>	<i>II</i>	<i>Vikar AJ</i>	<i>AJ420E</i>	<i>085A0113</i>			
	<i>200 g</i>	<i>100 mg</i>	<i>5 mg</i>	<i>II</i>	<i>Goldtech</i>	<i>GTJ</i>	<i>50927749</i>			
	<i>200 g</i>	<i>200 mg</i>	<i>10 mg</i>	<i>II</i>	<i>Wensar</i>	<i>PGB301</i>	<i>45223</i>			
	<i>1 kg</i>	<i>2 g</i>	<i>100 mg</i>	<i>II</i>	<i>Glory</i>	<i>GRYT</i>	<i>150200846</i>			
	<i>2 Nos. 600 g</i>	<i>200 mg</i>	<i>10 mg</i>	<i>II</i>	<i>GTEP Goldtech</i>	<i>GTJ</i>	<i>12043193</i>			
	<i>20 kg</i>	<i>5 g</i>	<i>0.1 g</i>	<i>D</i>	<i>Wensar</i>	<i>TT020H</i>	<i>38862</i>			
	<i>30 kg</i>	<i>100 g</i>	<i>2 g</i>	<i>II</i>	<i>Goldtech</i>	<i>GTA30K</i>	<i>40311521</i>			

Total Rs. *4890.00/- (part of 16125.00/-)* realized vide T. Receipt/Money receipt No. *GRN: 19202223026632641* dated *24/01/2023*

Repaired by / Used by *Electronics* (Signature)

Next verification due on *21/02/2024*

Note: In the case of rejected weights, measures, the Legal Metrology Officer shall give separate Certificate of rejection mentioning the reasons of rejection against each item.

*As per Rule - 22 : Certificate of Verification is to be exhibited in a conspicuous place.

[Signature]
 Legal Metrology Officer

Inspector of Legal Metrology
 Unit No : 67, Kolkata (South)
 Govt. of W.B.



[See Rule 15(3)]
 Government of West Bengal
OFFICE OF THE CONTROLLER OF LEGAL METROLOGY
Certificate of verification

SBVC A **1922931** ORIG

Name of Legal Metrology Officer: **R. PAUL, WB-262** Unit No.: 67, Kolkata (South)
 No.

I hereby certify that I have this day ^{re}verified and stamped / [✓]rejected the under mentioned weights, measures, etc.
 Belonging to **Deys Medical Stores MFG. Ltd.** Locality **Gr. Bondel Road, PO-19, P.S. - Gariahat**

Quantity	Denomination		Weighing instrument				Measuring instrument	Verification Fee		Carriage Conveyance adjusting ch etc.
	Weights	Measures	Capacity	Class	Manufacturer	Type		Rs.	P.	
1	2	3	4	5	6	7	8	9	10	
WI.	6 NOS.	Max: 200kg, Min: 500g, e=10g		II	Gold Tech	Model - GTP.				
16.	4 NOS.	Max: 500g, Min: 400mg, e=20mg	SL No. - 13057327, 21250256, 21230257, 13057326, 21037891, 21037893	II	Gold Tech.	Model - GTJ.		8550/-	w/n 16/	
	2 NOS.	Max: 600g, Min: 200mg, e=10mg	SL No. - 2004/194, 91143423, 30511515, 91143428,	II	Wensa	Model - PGB-600.				
	2 NOS.	Max: 60kg, Min: 250g, e=5g	SL: 63590, 87435	II	Goldtech	Model - GTP.				
	2 NOS.	Max: 30kg, Min: 30g, e=1g	SL: 6773, 21145325	II	Goldtech	Model - GTA.	SL-19A310365, 2014033,			

Total Rs. **8550/- (part of 16125/-)** realized vide **Receipt/Money receipt No. GIN: 192022230266532641** dated **24/01/2023**

Repaired by / Used by **Deys Medical Stores** (Signature)
 Next verification due on **21/02/2024**

Note : In the case of rejected weights, measures, the Legal Metrology Officer shall give separate Certificate of rejection mentioning the reasons of rejection against each item.
 *As per Rule - 22 : Certificate of Verification is to be exhibited in a conspicuous place.

R. Paul
 21/02/2023
 Legal Metrology Officer

Inspector of Legal Metrology
 Unit No : 67, Kolkata (South)
 Govt. of W.B.



[See Rule 15(3)]
 Government of West Bengal
OFFICE OF THE CONTROLLER OF LEGAL METROLOGY
 Certificate of verification

ORIG
 SBVC A **1922930**

Name of Legal Metrology Officer: **R. PAUL, WB-262** Unit No.: 67, Kolkata (South)
 No.

I hereby certify that I have this day ^{re} verified and stamped / [✓] rejected the under mentioned weights, measures, etc.

Belonging to **Deys Medical Stores (MFG.) Ltd.** Locality **62, Dandel Road, 401-12, P. S. - Gariahat**

Quantity	Denomination		Weighing instrument				Measuring instrument	Verification Fee		Carriage Conveyance adjusting ch etc.
	Weights	Measures	Capacity	Class	Manufacturer	Type		Rs.	P.	
1	2	3	4	5	6	7	8	9	10	
WT. 8 C/S 3	2 Nos. 2 Nos. Counter Machine, 10kg, 5kg, 1kg	Max: 100kg	d = 100g	III	Avery	Dial type	Model - IND3205 CLE.	2685/- 2685/-	As per rule	
		Max: 100kg	d = 200g	III	Avery	Dial type	Model - IND3303CLS.			
		Max: 500kg	d = 1kg	III	Avery	Stedjard	Model - IND321800A.			
		Max: 1000kg	d = 2kg	III	Avery	Dial type	Model - 3218CLE.			
		Max: 200kg	d = 500g	III	Avery	Model - 3205CLE, 21L - C8412403.	Model - 3218CLE			
		Max: 50kg	d = 100g	III	Avery	Model - 3309CFS, 21L - C846914.	Model - 3218CLE			

GRN: 192022230266532641

Total Rs. **2685/-** realized vide T-Receipt/Money receipt No. dated **24/01/2023**

Repaired by / Used by **Dep 21/02/2023**
 For HINDUSTHAN WEIGHING MACHINE & CO.
 (Signature)

Next verification due on **21/02/2024**

R. Paul
 21/02/23
 Legal Metrology Officer

Note: In the case of rejected weights, measures, the Legal Metrology Officer shall give separate Certificate of rejection mentioning the reasons of rejection against each item.

*As per Rule - 22: Certificate of Verification is to be exhibited in a conspicuous place.